



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Interface Performance Materials, Inc. / Fulton Plant
2885 State Route 481, Fulton, NY 13069

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

Mechanical Testing
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

May 5, 2005

Issue Date:

September 30, 2015

Expiration Date:

January 31, 2018

Accreditation No.:

59339

Certificate No.:

L15-313

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Interface Performance Materials, Inc./ Fulton Plant

2885 Ste Route 481, Fulton, NY 13069
 Contact: Mike Osborne Phone: 315-592-8104

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical ^F	Non-Metallic Gasket Materials	Thickness	ISI 10-021-039 ASTM F104 Standard Classification	0.002 5 mm
		Basis Weight	ISI 10-021-012 ASTM F104 Standard Classification	1 g/m ²
		Density	ISI 10-021-012 ASTM F1315 Standard Classification	0.01 g/m ²
		Compressibility and Recovery	ISI 10-021-028 ASTM F36	0.1 %
		Sealability	ISI 10-021-029 (Method B) ASTM F37	0.1 mL/hr
		Creep Relaxation	ISI 10-021-030 (Method B) ASTM F38	0.1 %
		Fluid Resistance	ISI 10-021-027 ASTM F146	0.1 %
		Flexibility	ISI 10-021-032 ASTM F147	Pass/Fail
		Tension Testing	ISI 10-021-017 ASTM F152	0.01 mPa
		Adhesion to Metal Surfaces	ISI 10-021-031 ASTM F607	0.1 lbf
		Dimensional Stability	ISI 10-021-033 ASTM F1087	0.1 %
Compressive Strength and Elevated Temperatures (Crush –Extrusion)	ISI 10-021-035 ASTM F1574 (Sec 9.3)	0.1 %		

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer ^F would mean that the laboratory performs this testing at its fixed location.